

Identification of the product

Product name	R-DiaLeg		
Product reference	DDGR-50-L100	Production date	2018-07
Lot number	05008G	Expiry date	2020-07



Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	29.5
Positive Control (Ct/Cp) :	YD	30.0 ± 3.3	29.9
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function :	Signature:	Date:
 Dominique d'Amice QA Officer		26/10/2018

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.