

Identification of the product

Product name	S-DiaGono		
Product reference	DDGS-40-L100	Production date	2019-10
Lot number	2019J07/II	Expiry date	2021-10


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	27.0 ± 3.3	28.7
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function	Signature:	Date:
<div style="border: 1px solid black; padding: 2px; display: inline-block;"> Dominique d'Amico QA Officer </div>		04 -11- 2019

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.