

Identification of the product

Product Name	H-DiaHSV		
Catalog Number	DDGH-40-L100	Production Date	2016-05
Lot Number	2016E27/I	Expiry Date	2017-10

Analysis procedure

Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	OPTIMA Multiplex 2x DNA
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Specifications	Results
Positive Control (Ct/Cp) HSV 1	27.4 ± 3.3	FAM: 26.7
Positive Control (Ct/Cp) HSV 2	26.6 ± 3.3	YD: 26.6
Negative Control (Ct/Cp)	Negative	Negative

Approval

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: D. KUSINDA QA OFFICER Date: 08/06/2016

