

Identification of the product

Product Name	Dientamoeba fragilis real-time PCR kit		
Catalog Number	DIA-DF-050	Production Date	2015-02
Lot Number	06005B	Expiry Date	2017-02

Analysis procedure

Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	QuantiFast Multiplex PCR+R kit (Qiagen)
Real Time PCR equipment (Supplier)	LightCycler 480 II(Roche)

QC Results by Real Time PCR

Testing	Specifications	Results
Positive Control (Ct/Cp)	FAM : 30.0 ± 3.3	FAM : 30.0
Negative Control (Ct/Cp)	Negative	Negative

Approval

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: F. Naye QA office

Date: 20 Apr 2015

