

Identification of the product

Product Name	M-DiaSero		
Catalogue Number	DDGM-20-L040	Production Date	2016-01
Lot Number	2016J21/I	Expiry Date	2018-01

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	25.0
Master Mix reference	OPTIMA Multiplex 2x DNA
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) N. tot :	FAM	30.0 ± 3.3	30.2
Positive Control (Ct/Cp) N. gB :	FAM	30.0 ± 3.3	31.4
Positive Control (Ct/Cp) N. gC :	FAM	30.0 ± 3.3	29.6
Positive Control (Ct/Cp) N. gA :	FAM	30.0 ± 3.3	31.2
Positive Control (Ct/Cp) N. gW135 :	FAM	30.0 ± 3.3	29.2
Positive Control (Ct/Cp) N. gY :	FAM	30.0 ± 3.3	30.3
Negative Control (Ct/Cp)	FAM	Negative	Negative

Approval

Name (function) : A. ZALEWSKI QA/QC Manager	Signature: 	Date: 10 NOV 2016
--	---	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.