

Identification of the product

Product Name	O-DiaADV		
Catalog Number	DDGO-10-L100	Production Date	2016-06
Lot Number	01006F	Expiry Date	2018-06

Analysis procedure


Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	Universal Master Mix UNG (Diagenode)
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Specifications	Results
Positive Control (Ct/Cp)	30.0 ± 3.3	FAM : 28.4
Negative Control (Ct/Cp)	Negative	Negative

Approval

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: R. ZACEDALS  Date: 08 Aug 2016
QA/PCR manager