

**Identification of the product**

Product Name	O-DiaENT		
Catalog Number	DDGO-30-L100	Production Date	2015-11
Lot Number	01005K	Expiry Date	2017-11

**Analysis procedure**


Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	TaqMan Fast Virus 1-Step Master Mix (Applied Biosystems)
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC Results by Real Time PCR**

Testing	Specifications	Results
Positive Control (Ct/Cp) Enterovirus	30.0 ± 3.3	FAM : 31.4
Negative Control (Ct/Cp)	Negative	Negative

**Approval**

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: A. ZACARIS  Date: 15 Feb 2016  
*PA / QC Manager*